





# Calibration ID: Sample EO16-851

Camera Information			Calibration Summary	
Part Information	Order Number		<b>Intrinsic Parameters</b> EFLx EFLy: [30339.6 30339.6] pixels PPx PPy: [2128.4 1558.7] pixels Radial Distortion Coeff: [1.5926 10.5372 146.7267] Tangential Distortion Coeff: [0.0026 0.0075] Tangential Distortion Poly Coeff: []	
	Order Date			
	Revision	A		
	Calibration Date	09-Apr-2024 18:46:12		
Sensor	Part Number	NANO-M4020	<b>Extrinsic Parameters</b> Rx Ry Rz: [-1.5738 0.0015 -1.5778] rad Euler XYZ Intrinsic x y z: [-0.01376 -0.01893 -0.08753] m	
	Serial Number	H2348628		
	Sensor Size	4112 x 3008		
Lens	Part Number	16851		
	Serial Number	0001		
Intrinsic Summary	Reprojection Error [pix]	0.5444		
	Model Type	BrownConradyModel		



For more information on how to deploy this calibration, visit: [www.quartus.com/pixeltraq](http://www.quartus.com/pixeltraq) or scan

