





Calibration ID: Sample EO16-852

Camera Information			Calibration Summary	
Part Information	Order Number		Intrinsic Parameters EFLx EFLy: [46817.1 46816.0] pixels PPx PPy: [2097.7 1635.3] pixels Radial Distortion Coeff: [2.0840 80.4720 -16911.6000] Tangential Distortion Coeff: [0.0047 0.0053] Tangential Distortion Poly Coeff: []	
	Order Date			
	Revision	A		
	Date Recorded	09-Apr-2024 19:11:12		
Sensor	Part Number	NANO-M4020	Extrinsic Parameters Rx Ry Rz: [-1.5725 0.0015 -1.5763] rad Euler XYZ Intrinsic x y z: [-0.01183 -0.01767 -0.16477] m	
	Serial Number	H2348628		
	Sensor Size	4112 x 3008		
Lens	Part Number	16852		
	Serial Number	0001		
Intrinsic Summary	Reprojection Error [pix]	0.6499		
	Model Type	BrownConradyModel		



For more information on how to deploy this calibration, visit: www.quartus.com/pixeltraq or scan

